Republic of the Philippines

**CITY OF NAGA**

J. Miranda Ave., Concepcion Pequeña, Naga City 4400

**BIDS AND AWARDS COMMITTEE (BAC)**

# BID BULLETIN NO. 1

29 November 2022

This Bid Bulletin No. 1 is issued to modify or amend items in the Bidding Documents of below indicated procurement, to wit:

| **Contract Ref. No.** | **Name of Project** |
|---|---|
| 2022-041 | On Item No. 1: Supply and Delivery of Oto-Acoustic Emission (OAE) TEOAE/ DPOAE without label printer (1 Lot)<br><br>**Remove: Without Label Printer**<br><br>On Item No. 2: The following are the minimum technical specifications for the 2 Units **ADVANCED FETAL/MATERNAL MONITOR**<br><br>**Standard Feature**<br>• 8.5" Colour graphic touchscreen display<br>• Integral 150mm Printer<br>• Dawes-Redman 100K CTG Analysis<br>• Labour TREND function<br>• Twins capable<br>• Fetal ECG (choice of electrode type)<br>• Maternal ECG<br>**Interfaces**<br>• RS232<br>• Ethernet (TCP/IP)<br>• USB (x2)<br>• Wireless System port<br>**Accessories**<br>• Ultrasound transducer<br>• Contractions transducer<br>• Patient Event Marker<br>• FECG leg plate interface cable (choice of electrode type)<br>• MECG interface cable<br>• IntraUterine Pressure interface cable<br>• Sonicaid Freedom - wireless transducer system<br>• Mobile trolley with storage trays |

| | |
|---|---|
| | • Wall mount<br>**<u>Factory Fit Options</u>**<br>• Maternal Blood Pressure (NiBP + MHR)(inc 1x med, 1x large cuff)<br>• Maternal SpO2 (O2 sat + MHR)(choice of sensor type)<br>• Triplets capable<br>• eCTG (no printer model)<br>• Integral re-chargeable battery<br>• DVI video output port<br>• Paper tray insert for pre-printed paper (choice of paper types)<br><br>Others: Preference to supplier(s) with local branch/service center in Metro Naga or Bicol Region. |

For guidance and information of all concerned.

**ENGR. ALEXANDER N. CANING**
BAC Chairperson

PhilGEPS, Bulletin boards, LGU website