



Republic of the Philippines
CITY OF NAGA
J. Miranda Ave., Concepcion Pequeña, Naga City 4400
BIDS AND AWARDS COMMITTEE (BAC)



BID BULLETIN NO. 1

December 22, 2022

This Bid Bulletin No. 1 is issued to modify or amend items in the Bidding Documents of below indicated projects, to wit:

Contract Ref. No.	Name of Project
CRN 2022-041	Supply and Delivery of Oto-Acoustic Emission (OAE) TEOAE/ DPOAE without label printer (a republication after a failed bidding)

Amendments made to the technical specifications including conditions added as an essential part of the bid document, to wit:

ON OTO-ACOUSTIC EMISSION (OAE) TEOAE/DPOAE


Specifications	Amendment
Range: DPOAE Quick/Diagnostic Test Modes: 800 HZ to 10 hz, up to 30 points per octave high resolution	Removed
Stimulus levels from 20 dB HL to 6.5 DB HL	30 dB SPL to 65 dB SPL
Frequency range: 1.0 kHz to 8 kHz	OK
TEOAE Quick/Diagnosis Test modes: 07 to 4 Khz Quick Test/ Diagnostic test with octave bands with multiple SNR settings and criteria	Removed
Capable of both screening and diagnostic	Upgradable got screening and diagnostic
Full color touchscreen operating	OK
Capable of data transmission to a centralized server via internet/mobile data	Removed
Upgraded of DP threshold estimation	Upgradable DP threshold estimation
With second probe for simultaneous testing of both earths	OK
Frequency Range: 1.0 kHz to 8 kHz	Removed
With instrument charger, accessory	OK

box and carrying bag	
Portable type	OK
Can store up to 100 tests	OK
Interchangeable OAE probe	OK
Capable of DP threshold estimation	Removed
10 hours battery operating hours	Minimum 6 hours battery operating hours
Stimulus levels 20 db hl to 65 hl	Removed
Accessories	
1 pc. OAE cable	OK
40 pcs Pediatric earprobe (4 types)	40 pcs. Pediatric Eartips (4 types)
40 pcs adult earprobe (4 types)	40 pcs. Adult Eartips (4 types)
5 pcs ear tips pedia	5 pcs Earprobe Tips Pedia
5 pcs ear tips adult	5 pcs Earprobe Tips Adult
1 pc cleaner	OK
1 set charger with USB cord	OK
2 pcs Rechargeable battery with at least 8 hours battery life auto-volt	Removed
1 pc carrying bag	OK

Other conditions:

1. Three (3) years warranty, free calibration.
2. Must provide temporary replacement while the probe is being calibrated offsite.
3. Equipment must have a NHSRC Certificate and FDA CMDN Certificate to be attached as part of the technical documents.
4. The equipment brand is being generally used by government hospitals in Camarines Sur.

For guidance and information of all concerned.


ENGR. ALEXANDER N. CANING
 BAC Chairperson